

Total Lift Bed

Tilting Instructions

1. Remove all extra pillows and ensure there are no obstructions from lines or leads. Pause tube feeds and wear appropriate socks or footwear. Ensure the bed is not touching the wall.
2. Position straps below the knee and above or across the knee and chest, being cautious of lines. Different strap setups are available; seek guidance from a super user for assistance.
3. To activate the tilt function, press and hold the bottom two buttons (11 & 12) together until a beep confirms tilt mode unlock.
4. Press and hold the bottom right button to initiate tilting (12).
5. Align the patient's feet/knees properly as the footplate moves towards them.
6. The bed will gradually tilt up; monitor for line clearance using the inclinometer on the right side under the top bed rail to gauge the tilt angle.
7. Tilt to 10 degrees less than what therapy services recommend, then evaluate the patient's response. Document the start time of therapy and total duration on the chart.
8. If the patient tolerates well, increase the tilt degree based on specified parameters.
9. To end the session, hold down the CPR flat button until the patient is back to zero degrees.
10. Adjust the patient's position as needed, ensuring the mattress settings for low air loss and alternating pressure are suitable for the patient's requirements.



Day	Time Start	Time Stop	Degree Achieved	Weight Bearing	Initial/Job Title

